







Stationary operation of CoMo-system in active wall station



CoMo with connected detector for dose rate measurement



CoMo placed in floor control device

Efficiencies for various radionuclides

Average values from measurements  
with 100 cm<sup>2</sup> compound

C-14 approx. 14 %

F-18 approx. 18 %

P-32 approx. 25 %

S-35 approx. 12 %

Cl-36 approx. 42 %

K-40 approx. 30 %

Co-57 approx. 7 %

Co-60 approx. 27 %

Sr-89 approx. 27 %

Sr-90 / Y-90

(related to Sr-90) approx. 42 %

Tc-99m approx. 3 %

In-111 approx. 8 %

I-123 approx. 7 %

I-125 approx. 12 %

I-131 approx. 21 %

Cs-137 approx. 35 %

Au-198 approx. 23 %

Ti-204 approx. 43 %

Am-241 á approx. 22 %

P-238 á approx. 12 %

U-238 á approx. 26 %

---

© Elimpex-Medizintechnik, Spechtgasse 32, A-2340 Moedling, Austria  
phone +43-2236-410450  
fax +43-2236-410459

---

